

Certificate of Analysis

Catalog #: UVTM00I
Description: Universal COVID-19 Viral Transport Medium
Lot #: UBP042820
Expiration Date: 04/28/2021
Country of Manufacture: USA
Storage: 2-8°C


Notes: Our products are prepared by a validated aseptic sterile filtration process to ensure that all products comply with the industry's Sterility Assurance Level (SAL) of at least 10⁻³ (i.e. demonstrating that a product produced in this manner has no more than 1 random contaminant per 1000 units).

This product was tested in accordance with the CDC currently approved guidelines and Neuromics specifications and procedures. Each lot is tested on Blood Agar to determine sterility of the final product. Testing procedures are maintained in compliance with the current versions of the USP and/or EP, where applicable.

For Research or Further Manufacturing Use Only Not for use in Diagnostic Procedures. The product is for transporting specimens only. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.

TEST PARAMETERS	SPECIFICATION	RESULT
pH	7.30 4 ₋ /-0.2	7.1
Appearance (Color)	Clear	Pass
Osmolarity	285 +/-5	281 mOsm/kg
Endotoxin	≤0.25 EU/mL	< 0.0010 EI- J/mL
Mycoplasma	Tested Negative	Negative
Sterility	Pass	Pass

For further information Email: rose@neuromics.com. Following signatures indicate the above material has met all quality specifications and has been reviewed by a Quality representative.



Quality Control