

Certificate of Analysis

Catalog #:	VTM001
Description:	Universal COVID-19 Viral Transport Media
Lot #:	UBP042820
Expiration Date:	04/28/2021
Country of Manufacture:	USA
Storage:	2-8°C. Store in Dark

TEST	SPECIFICATION	RESULT
pH	7.30 4 +/-0.2	7.1
Appearance (Color)	Clear	Pass
Osmolarity	285 +/-5	281 mOsm/kg
Endotoxin	≤0.25 EU/mL	< 0.0010 EI-J/mL
Mycoplasma	Tested Negative	Negative
Sterility	Pass	Pass

Notes:

This product was tested in accordance with the CDC currently approved guidelines (SOP#: DSR-052-02) and Neuromics specifications and procedures. Each lot is tested on Blood Agar to determine sterility of the final product.

Virus Transport Media (VTM) is designed for the transport of viruses for research and testing purposes. This media is not to be taken internally and is for the transport of specimens only.



 Quality Control